

Alcan Technology & Management Ltd.
 Badische Bahnhofstrasse 16
 CH-8212 Neuhausen Switzerland



Certificate of Analysis
Certified Reference Material
Aluminium Base

442/03

Date of certification: 30-APR-1991/03-JUL-1995

Element	Si	Fe	Cu	Mn	Mg	Cr	Ni	Zn	Ti	Ag	As	B	Ba
Weight % (w/w)	8.06	0.58	2.37	0.25	0.098	0.010	0.050	0.71	0.117				
Tolerance ±	0.15	0.02	0.05	0.02	0.005	0.002	0.003	0.02	0.005				

Element	Be	Bi	Ca	Cd	Ce	Co	Ga	Hg	In	La	Li	Mo	Na
Weight % (w/w)			0.0070										
Tolerance ±			0.0005										

Element	P	Pb	Sb	Sc	Se	Sn	Sr	Ta	Tl	V	W	Zr
Weight % (w/w)	0.0007	0.11				0.087				0.013		0.017
Tolerance ±	0.0002	0.01				0.007				0.001		0.002

We hereby certify the above analysis

Franz Hochleitner

* Results in µg/g. Exact values certified for S-samples only.
 * Angaben in µg/g. Genauere Werte nur für S-Proben zertifiziert.